

ASTM D 1665 - D941

SCOPE

This test method covers the determination of specific viscosity of tars and their fluids products. It does not determine absolute viscosity, but is an empirical flow test. Only by conforming strictly to requirements of the test method are reproducible results obtained.

ENGLER VISCOMETER FOR BITUMEN - REF 27235-

MAIN CHARACTERISTICS

- ✓ Place Engler viscometer
- ✓ Electric heating
- ✓ Temperature digital display
- ✓ Temperature from room up to +99.9°C
- ✓ Polished brass test cup with stainless steel efflux tube
- ✓ Water bath motor stirrer 2800 rpm
- ✓ Safety thermostat
- ✓ Cooling coil

Apparatus delivered without accessories

SCOPE OF DELIVERY

REF 27235 1 Place Engler viscometer

NECESSARY ACCESSORIES

REF 521155 Engler flask, 50 ml
REF 27251 Engler thermometer with ring +10/+55°C/0.5°C
REF 27252 Engler thermometer with ring +10/+110°C/1°C
REF 27253 Engler flask 100+100 ml
REF 27254 Engler flask 200+40 ml

OPTIONAL ACCESSORIES

REF 22236 Cooling Cryostat (-20/+100°C ±.0.03)



ORDERING INFORMATION

REF 27235

Engler viscometer For use on AC 230 V, 50Hz – 2A

REF 27235 115V on request

(W) 330x (D) 270x (H) 540 mm (± 8kg)

CONTACT : sales@normalab.com

NORMALAB FRANCE SAS
ZA Caux Multipôles 1 - F-76 190 Valliquerville
Tel. : +33 232 700 100
Fax : +33 232 704 732

www.normalab.com

REF 27235 Leaflet 15

DISTRIBUTED BY

G-Labo Germany
Bgm.-Horneffstr.26
69509 Mörlenbach
Tel.: + 49 6209 797100
Fax: + 49 6209 797101
Mail: info@g-labo.de
Web: www.g-labo.de