

ASTM D 70 - ISO 3828 - IP 189/IP 190**SCOPE**

Standard test method for density of semi-solid bituminous materials.

DENSITY OF SEMI-SOLID BITUMINOUS MATERIALS
PYCNO METER METHOD
- REF 24615 -**MAIN CHARACTERISTICS**

Thermostatic bath TC 16:

- ✓ Temperature range 5°C above ambient temperature to +250°C
- ✓ Capacity: 16L
- ✓ Electronic regulation
- ✓ Stability: +/- 0,01°C
- ✓ Cooling coil
- ✓ Heating: 1200 W
- ✓ Pump capacity: 7,5 L/min
- ✓ Opening 180x210 mm
- ✓ Depth: 200 mm

NECESSARY ACCESSORY

REF 11532 ASTM thermometer (63 C)

STANDARDIZED GLASSWARE

REF 20847 Pycnometer A (24/30 MI)
REF 23229 Pycnometer C (24/30 MI)
REF 23230 Pycnometer D (24/30 MI)
REF 24624 Pycnometer B (24/30 MI)

**ORDERING INFORMATION**

REF 24615

For use on 230 V-50/60 Hz – 1600 W

CONTACT : sales@normalab.com

NORMALAB FRANCE SAS
14 rue des Lilas - F-76 210 Lintot
Tel. : +332 35 38 59 59
Fax : +332 35 38 78 55

www.normalab.com

PYCNO METER METHOD Brochure 14

DISTRIBUTED BY

G-Labo Germany
Bgm.-Horneffstr.26
69509 Mörlenbach
Tel.: + 49 6209 797100
Fax: + 49 6209 797101
Mail: info@g-labo.de
Web: www.g-labo.de